

Declaration of Conformity

Identification of the product:	Infrared Gas Analyzer	
Name and address of the manufacturer	Name and address of the authorized representative:	
Fuji Electric Co., Ltd. 1, Fuji-machi, Hino-shi, Tokyo 191-8502 JAPAN	FUJI Electric France SAS 46 rue Georges Besse ZI Brezet 63039 Clermont-Ferrand Cedex 2 FRANCE	
This declaration of conformity is issued under the sole responsibility of the manufacturer.		
Object of the declaration:	Infrared Gas Analyzer (ZPA , ZPB , ZPG series)	
The object of the declaration described above is in conformity with the relevant EU harmonization legislation:		
2014/30/EU	Electromagnetic compatibility	(EMC)
2014/35/EU	Low Voltage	(LVD)
2011/65/EU	Restriction of the use of certain hazardous substances	(RoHS)
References to the relevant harmonized standards used or references to the technical specifications in relation to which conformity is declared:		
EN 61326-1: 2013 (Table 2 Industrial environment) (Class A Group 1)	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements	
EN 61326-2-3: 2013 (Table 101)	Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning	
EN55011:2009 , A1:2010	Industrial, scientific and medical equipment. Radio-frequency disturbance characteristics. Limits and methods of measurement	
EN61000-3-2:2014	Electromagnetic compatibility (EMC). Limits. Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)	
EN61000-3-3:2013	Electromagnetic compatibility (EMC). Limits. Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection	
EN61010-1:2010	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements	
EN62311:2008	Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz - 300 GHz)	
EN 50581: 2012	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances	
Additional information:		
Year of CE marking(LVD):	16(last two digits of year)	
Technical Documentation No	TN5A2782a	
Signed for and on behalf of the above named manufacturer:		
Place and date of issue:	Hino-shi, Tokyo JAPAN 11-JUN-2021	
Name, function:	KAZUNAO YONEYAMA Manager, Quality Assurance Dept. Tokyo Factory Fuji Electric Co., Ltd.	
Signature:	<i>K. Yoneyama</i>	