

UKCA Declaration of Conformity

Identification of the product: Zirconia Oxygen Analyzer

Name and address of the manufacturer
 Fuji Electric Co., Ltd.
 1, Fuji-machi, Hino-shi, Tokyo 191-8502
 JAPAN

Name and address of the authorized representative:
 FUJI Electric France SAS
 46 rue Georges Besse ZI Brezet
 63039 Clermont-Ferrand Cedex 2
 FRANCE

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Object of the declaration: Zirconia Oxygen Analyzer (ZFK7)

The object of the declaration described above is in conformity with the relevant UK harmonization legislation:

S.I. 2016 No.1091	The Electromagnetic Compatibility Regulations 2016	(EMC)
S.I. 2016 No.1101	The Electrical Equipment (Safety) Regulations 2016	(LVD)
S.I. 2012 No.3032	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012	(RoHS)

References to the relevant harmonized standards used or references to the technical specifications in relation to which conformity is declared:

EN 61326-1: 2013 (Table 2 Industrial environment) (Class A Group 1)	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61326-2-3: 2013 (Table 101)	Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning
EN55011:2009 , A1:2010	Industrial,scientific and medical equipment. Radio-frequency disturbance characteristics. Limits and methods of measurement
EN61000-3-2:2014	Electromagnetic compatibility (EMC). Limits. Limits for harmonic current emissions (equipment input current ≤ 16 A per phase)
EN61000-3-3:2013	Electromagnetic compatibility (EMC). Limits. Limitation of voltage changes, voltage fluctuations and flicker in public low-voltage supply systems, for equipment with rated current ≤ 16 A per phase and not subject to conditional connection
EN61010-1:2010,A1:2019	Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements
EN62311:2008	Assessment of electronic and electrical equipment related to human exposure restrictions for electromagnetic fields (0 Hz - 300 GHz)
EN IEC 63000:2018	Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances


Additional information:

Year of UKCA marking(LVD): 22 (last two digits of year)
 Technical Documentation No TN5A2737

Signed for and on behalf of the above named manufacturer:

Place and date of issue: Hino-shi, Tokyo JAPAN 30-Sep-2022
 Name, function: KAZUNAO YONEYAMA
 Manager, Quality Assurance Dept. Tokyo Factory
 Fuji Electric Co., Ltd.

Signature:



Doc. No. TN5A2737 (1/1)